



KERALA GAZETTE

കേരള ഗസറ്റ്

PUBLISHED BY AUTHORITY

ആധികാരികമായി പ്രസിദ്ധപ്പെടുത്തുന്നത്

Vol. II	}	THIRUVANANTHAPURAM, TUESDAY	31st December 2013	}	No.
വാല്യം 2		തിരുവനന്തപുരം, ചൊവ്വ	2013 ഡിസംബർ 31		നമ്പർ
			10th Pousha 1935		52
			1935 പൗഷം 10		

PART III

Stores Purchase

Medical Education Department

TENDER NOTICES

(1)

No. C1/22735/2013/MCT. 6th December 2013.

Sealed competitive tenders in the form prescribed are invited for the the supply of the following items at four Pharmacy Colleges attached to Government Medical Colleges of Thiruvananthapuram, Kottayam, Alappuzha & Kozhikode. Tenders should be superscribed as tender numbered and addressed to the Principal, Medical College, Thiruvananthapuram, Pin-695 011. Details can be had from this office on working hours till 3.30 p. m.

229/13	Disintegration Test Apparatus—4 Nos.	1,680	840
230/13	Microwave Digestor—4 Nos.	1,680	840
231/13	Digital Refractometer—4 Nos.	3,780	1,890
232/13	Digital Polarimeter—4 Nos.	4,410	2,205
233/13	Karl Fischer Apparatus—4 Nos.	1,260	630
234/13	Digital melting point Apparatus—4 Nos.	1,680	840
235/13	Multiparameter Tablet Machine—4 Nos.	2,520	1,260
236/13	Laboratory Filtration Kit—4 Nos.	1,050	525

Tender No.	Details of items with Quantity	Cost of tender form (VAT inclusive)	
		Original ₹	Duplicate ₹
228/13	Dissolution Test Apparatus—4 Nos.	6,300	3,150

237/13	Desk Top Computer—4 Nos.	420	210
238/13	Refrigerator (Specifications attached separately)—4 Nos.	420	210

Last date and time for sale of tender form—6-1-2014, 2.30 p. m.

Last date and time of receipt of tender form—7-1-2014, 2.30 p. m.

Opening date and time of tender—7-1-2014, 3 p. m.

Tender forms can be had from the Office the Principal, Medical College, Thiruvananthapuram-11 up to 3 p.m. on working days. Tender forms will be sent through post if required provided due amount with sufficient Postage Charge (₹ 50 each) is sent by way of Money Order in favour of the Principal, Medical College, Thiruvananthapuram, Kerala State, Pin-695 011. Money Order should reach this office on or before 21-12-2013. No other forms of remittance will be accepted.

തിരുവനന്തപുരം, കോട്ടയം, ആലപ്പുഴ, കോഴിക്കോട് മെഡിക്കൽ കോളേജുകളിലെ 4 ഫാർമസി കോളേജിലേക്ക് ആവശ്യമായ താഴെ പറയുന്ന സാധനങ്ങൾ വിതരണം ചെയ്യുന്നതിന് മുദ്രവച്ച ദർഘാസുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു. ദർഘാസുകൾ ദർഘാസ് നമ്പരോടുകൂടി നിശ്ചിത തീയതിക്കുമുമ്പ് പ്രിൻസിപ്പാൾ, മെഡിക്കൽ കോളേജ്, തിരുവനന്തപുരം-695 011 എന്ന വിലാസത്തിൽ എത്തിച്ചിരിക്കേണ്ടതാണ്.

ദർഘാസ് നമ്പർ	സാധനങ്ങളുടെ പേര്	ദർഘാസ് ഫോറത്തിന്റെ വില	
		അസ്സൽ	ശരിപ്പകർപ്പ്
		₹	₹
228/13	ഡിസൊല്യൂഷൻ ടെസ്റ്റ് അപ്പാററ്റസ്—4 എണ്ണം	6,300	3,150
229/13	ഡിസിന്റഗ്രേഷൻ ടെസ്റ്റ് അപ്പാററ്റസ്—4 എണ്ണം	1,680	840
230/13	മൈക്രോവേവ് ഡൈജസ്റ്റർ—4 എണ്ണം	1,680	840
231/13	ഡിജിറ്റൽ റീഫ്രിജറേറ്റർ മീറ്റർ—4 എണ്ണം	3,780	1,890
232/13	ഡിജിറ്റൽ പോലാരി മീറ്റർ—4 എണ്ണം	4,410	2,205
233/13	കാൽ ഫിഷർ അപ്പാററ്റസ്—4 എണ്ണം	1,260	630
234/13	ഡിജിറ്റൽ മെൽറ്റിങ്ങ് പോയിന്റ് അപ്പാററ്റസ്—4 എണ്ണം	1,680	840
235/13	മൾട്ടിപാരാമീറ്റർ ട്രാൻസ് മെഷീൻ—4 എണ്ണം	2,520	1,260

236/13	ലബോറട്ടറി ഫിൻഡറേഷൻ കിറ്റ്—4 എണ്ണം	1,050	525
237/13	ഡെസ്ക്ടോപ്പ് കമ്പ്യൂട്ടർ—4 എണ്ണം	420	210
238/13	റെഫ്രിജറേറ്റർ—4 എണ്ണം	420	210

ദർഘാസ് ഫോറം വിൽക്കുന്ന അവസാന തീയതിയും സമയവും—6-1-2014, 2.30 മണി.

ദർഘാസ് ഫോറം സ്വീകരിക്കുന്ന അവസാന തീയതിയും സമയവും—7-1-2014, 2.30 മണി.

ദർഘാസ് തുറക്കുന്ന തീയതിയും സമയവും—7-1-2014, 3 മണി.

ദർഘാസ് ഫോറങ്ങൾ തപാലിൽ വേണ്ടവർ അയയ്ക്കുന്നതിനുവേണ്ട തുക (₹ 50) അധികമായി അയയ്ക്കേണ്ടതാണ്. ദർഘാസ് ഫോറത്തിന്റെ വില ഓഫീസിൽ നേരിട്ടോ മണിയോർഡർ ആയോ സ്വീകരിക്കുന്നതാണ്. അല്ലാതെയുള്ള എല്ലാ അപേക്ഷകളും നിരസിക്കുന്നതാണ്. ഫോറത്തോടൊപ്പം മുദ്രപത്രങ്ങളോ മറ്റേതെങ്കിലും ഡോക്യുമെന്റുകളോ അയയ്ക്കുവാൻ ആവശ്യപ്പെട്ടുകൊണ്ടുള്ളതും മേൽവിലാസം വ്യക്തമല്ലാത്തതുമായ മണിയോർഡറുകൾ നിരസിക്കുന്നതാണ്. മണിയോർഡറുകൾ വഴി ഫോറങ്ങൾ ആവശ്യപ്പെട്ടുകൊണ്ടുള്ള അപേക്ഷകൾ ഈ ആഫീസിൽ 21-12-2013-നുമുമ്പ് കിട്ടിയിരിക്കണം. താമസിച്ചുകിട്ടുന്ന മണിയോർഡറുകൾ നിരസിക്കുന്നതാണ്. ദർഘാസ് ഫോറം കൈമാറാൻ പാടില്ലാത്തതും ഒരിക്കൽ അടച്ച തുക തിരിച്ചു കിട്ടുന്നതുമല്ല. താമസിച്ചു കിട്ടുന്ന ദർഘാസുകൾ യാതൊരു കാരണവശാലും സ്വീകരിക്കുന്നതല്ല. കൂടുതൽ വിവരങ്ങൾ പ്രവൃത്തി ദിവസങ്ങളിൽ ഈ ആഫീസിൽ നിന്നും ലഭിക്കുന്നതാണ്.

No. C1/22735/2013/GMCT

1. Dissolution Test Apparatus

- Eight stations dissolution tester.
- Should be suitable for Type 1,2,5,6 in the same system.
- Interchangeable paddle and basket is required for Easy operation.
- Speed range: Variable speed from 25 to 250 RPM
- Speed accuracy: $\pm 1\%$ Speed Selection should be available via. Touch screen
- Facility for Multiple language support with preventive maintenance and calibration alert facility
- Ambient Temperature range: 5 to 40°C
- Display back lighted LCD with touch Screen Display.
- Complies with IP, USP, Ph.Eur. and JP specifications
- Bath: Clear acrylic molded water bath with Vibration free water circulator built-in the System with vibration isolators.

- Water Bath Temp Range should be 5 to 55 Deg. C with Temp Accuracy of $< \pm 2$ Deg. C
- Evaporation loss should be minimal $< 2\%$ in 24 Hours
- Minimum 15 programmable sampling intervals
- Vibration free smooth motorized lift movement
- Should have individual-Temp. measurement in each vessel.
- The system should have facility to measure Temp automatically at the predefined time point and between Time point for precise control of measurement.
- The system should come with table dropping mechanism with facility to drop the tablet simultaneously, time based and when desired temp is achieved in all vessel
- The system should have facility to upgrade in future for bathless type, as UV on Line and LC
- Dissolution
- The system should come along with the printer having print life of more than 10 years.
- The software option based on client server can also be considered.
- Additional Terms and Condition
- Manufacturer should have ISO 9001 Certificate for quality standards.
- On site comprehensive training for lab staff and support services till customers satisfaction with the system.
- Installation testing: Supplier of the instrument must provide free installation, commission and testing
- Should comply with 21 CFR parts 11.
- User/Technical/Maintenance manuals to be supplied in English.
- List of equipments available for providing calibration and routine preventive maintenance support as per manufacturer service/maintenance manual.

- Mechanical qualification system with certificate should be PC based—1 set
- Compliance report to be submitted in a tabulated and point wise manner clearly mentioning the page/para number of original catalogue/data sheet. Any point, if not substantiated with authenticated catalogue/manual, will not be considered.

Accessories to be quoted along with the system

1. SS Shaft
2. Electropolished Paddle
3. 40 Mesh Basket
4. 1 set of accessories for Type 5 and 6.
5. Individual Temp sensor for each vessel
6. Sampling cannula with filters
7. Set of Aligning tools for calibration purpose.
8. Options of Verified components.

No. C1/22735 (1) 13/GMCT

2. **Disintegration Test Apparatus**

Features:

- Complies with USP, IP, Eur. Pharmacopoeias
- Disintegration time registration of individual tablets
- Basket park-out from beaker on test completion
- Test results and validation gets printed
- individual motor and timer with external probe for both the baskets
- Bath with flip-top cover for easy cleaning magnetically coupled basket for easy loading

Specifications:

- Water bath: Clear acrylic moulded water bath
- Beaker: 1000 ml glass beakers (2 Nos.)
- Baskets: 2 Nos. each carrying 6 test position (dimensions as per latest USP)
- Guided disc Perforated-cylindrical disc- 6 Nos. in each basket (dimensions as per latest USP)
- Display: 20 × 4 LCD
- Printer interface: Parallel port
- Motor: Single shaft 6V DC motor 10kg cm torque (2 Nos.)
- Heater: 400 watts, S. S. 304
- Illumination: White LED
- Stirrer motor: DC motor 12V

- O Water level sensor: Float switch type level sensor
- O Power supply: 230 V AC. 50 Hz, 490 VA
- O No. of strokes: 30 + 0.5 strokes/min
- O Stroke height: 55 mm +2 mm
- O Readable temperature range: 10° C to 45° C
- O Controlled temperature range: 15° C to 40° C
- O Temperature resolution : + 0.2° C.
- O Timer setting: 23hrs 59 min. 59 sec. (for both baskets)

Standard accessories

- | | |
|--------------------------------------|---------|
| (i) Heater, stirrer and level sensor | 1 No. |
| (ii) Acrylic moulded water bath | 1 No. |
| (iii) 1000 ml glass beaker | 2 Nos. |
| (iv) 6 tube magnetic basket | 2 Nos. |
| (v) Guided disc | 12 Nos. |
| (vi) External probe with clamp | 2 Nos. |
| (vii) Keyboard | 1 No. |
| (viii) 3 tube bolus basket | |
| (ix) 1 tube bolus basket | |
| (x) Magnetic guided disc | |
| (xi) Printer | |

5. Microwave Digestor

The system should be designed for laboratory use with all SS case and corrosion resistant PFA coating, pressure stable circular oven for homogeneous microwave distribution. It should have massive security lid, electronic lid closure with 3 micro-switches safety interlocks to prevent microwave emission with open lid. Built in exhaust system for permanent cavity ventilation should be provided. The system should have temperature and pressure monitoring devices. Internal diagnostics such as magnetron status, magnetron temp, oven temp transformer, cover lock etc., shall be provided. Should be provided with microprocessor based controller with touch screen and control softwares. The system should have 12 vessels per rotor. The rotor should be supplied with 12 caps without pressure control and 1 cap with pressure control. The cap should be made of TFM. The system shall be supplied with a turn table of 60ml volume at 40bar operating pressure.

Specifications:

- | | |
|-------------------------|---------------|
| Microwave Frequency | : 2450 MHz |
| Microwave Power | : 1450 W |
| Temp. Measurement range | : 80 to 300°C |

- | | |
|----------------------------|------------------------------------|
| Max. temp (continuous) | : 230°C |
| Max. temp. (peak) | : 260°C |
| Temp precision | : $\pm 1^\circ \text{C}$ at 200° C |
| Pressure measurement range | : 0 to 150 bar |
| Pressure precision | : ± 5 bar |
| Programme memory | : 200 programmes |
| Data memory | : 500 data files |
| Interface | : RS. 232, Ethernet and USB |

6. Digital Refractometer

- Should have LCD Display with touch screen facility
- Should have indications for Refractive Index (nD), Brix (%), Temp (degree C)
- Temperature controller : Built-in Peltier Thermo-module
- Light source : LED Lamp
- Measurement Range :

Refractive index (nD)	: from 1.32422 to 170000
Brix	: 0.00 to 100.00%
- Minimum Indication :

Refractive Index (nD)	: 0.00001
Brix	: 0.01% (0.1%)
- Temp : 0.01°C
- Measurement Accuracy :

Refractive Index (nD)	: ± 0.0001
Brix	: $\pm 0.1\%$
Temp.	: $\pm 0.05^\circ \text{C}$
- Resolution :

Refractive Index (nD)	: 0.00001
Brix	: 0.01% (0.1%)
Temp	: 0.01°C
- Measurement Temperature: 5.00 to 75.00°C
- Temperature accuracy : $\pm 0.01^\circ \text{C}$
- Ambient Temperature : 10 to 40°C
- Output : RS-232 port, USB
- Printer : Dot matrix printer (ordinary Paper): 1 No.
- Self-Diagnosis Scale : The instrument can detect irregularities with intensity of light which can ensure accurate measurement
- Measurement Range :

Refractive Index (nD)	: from 1.32422 to 1.70000
Brix	: 0.00 to 100.00%
- Minimum Indication :

Refractive Index (nD)	: 0.00001
Brix	: 0.01% (0.1%)
Temp	: 0.01°C

- Measurement Accuracy :
Refractive Index (nD) : ± 0.0001
Brix : $\pm 0.1\%$
Temp : $\pm 0.05^\circ\text{C}$
- Resolution :
Refractive Index (nD) : 0.00001
Brix : 0.01% (0.1%)
Temp : 0.01°C
- Measurement Temperature: 5.00 to 75.00°C
- Temperature accuracy : $\pm 0.01^\circ\text{C}$
- Ambient Temperature : 10 to 40°C
- Output : RS-232 port, USB
- Printer : Dot matrix printer (ordinary Paper): 1 No.
- Self-Diagnosis Scale : The instrument can detect irregularities with intensity of light which can ensure accurate measurement

7. Digital Polarimeter

- Measurement Scales : Angle of Rotation, International Sugar Scale (without temperature compensation), International Sugar Scale (with Automatic Temperature Compensation), Specific Rotation, Concentration, Purity and Angle of Rotation (Temperature Compensation of Quartz Plate)
- Measurement Range : Angle of Rotation -89.999 to + 90.000° or -360 to +360°
International Sugar Scale -259 to + 259°
- Resolution : Angle of Rotation 0.001°
International Sugar Scale 0.001°Z
- Measurement Accuracy : Angle of Rotation Displayed value $\pm 0.005^\circ$ —(-45 to + 45°)
International Sugar Scale Displayed value $\pm 0.015^\circ\text{Z}$ (-130 to + 130°Z) (with a Standard Quartz plate)
- Repeatability : $\pm 0.003^\circ$ Angle of Rotation and $\pm 0.009^\circ\text{Z}$
Sensitivity up to OD2*
*OD = optical density : a measure of light absorbance.
OD2 = 1/100 attenuation and 1% transmission
- Temperature compensation range for International Sugar Scale : 10
- Display Panel : 7.5 inch color LCD (touch screen)
Measurement wavelength : 589 nm (D-line) with Cylindrical glass cell, 100mm and 200mm

8. Karl Fischer Titrator

- Volumetric Karl Fischer Titrator

- Should be able to detect 10ppm to 100% water content :
- Resolution : 0.01ml
- Mode : Blank, Titer Sample
- Results units : %w/w, % w/v, ppm, mg, mg/ml
- Real time clock : For Date & Time of Titration
- Report : Result report on display /Document on printer & PC

10. Digital Melting Point Apparatus

The video camera equipped melting point apparatus for the automatic determination of melting and boiling points. The apparatus shall have a real time video display in 6 x magnification. Should be compliant with Pharmacopeia methods including PH. EUR. 6.1.2.2.60, USP XXI 741, IP and JP and with GLP/GMP compliant protocols.

Specifications

Manual melting point determination	: Manual
Manual boiling point determination	: Manual
Automatic melting point determination	: Automatic
Automatic boiling point determination	: Automatic
Positions for melting point capillaries	: 3
Positions for boiling point capillaries	: 1
Storable methods for melting point	: 50
Storable methods for boiling point	: 50
Precision magnifying lens	: Yes
Magnification of lens	: 2.5 ×
Digital camera	: Yes
Video function	: Yes
Video run time	: 350 min at 1°C/min, 700 min at 0.5°C/min
Magnification, Display	: 6 ×
Display Colour,	: TFT, 320 × 240, 3.5"
Determination temperature range	: Ambient + 10°C to 400°C
Temperature resolution	: 0.1°C

Accuracy melting point at 0.5°C/min	:	
up to 250°C	:	+/- 0,3°C
250 to 400°C	:	+/- 0,3°C to +/-0,5°C
Repeatability melting point at 0.5°C/min	:	+/- 0,1°C
Accuracy boiling point at 1.0°C/min up to 400°C	:	+/- 0,5°C
Repeatability boiling point at 1.0°C/min	:	+/- 0,3°C
Temperature gradients, °C/min	:	0.1,0.2,0.5,1,2,5,10,20
Heat-up time (50°C-350°C)	:	~ 4 min
Cool-down time (350°C-50°C)	:	~ 13 min
Electrical supply	:	100-240 (+/- 10%), 50-60Hz

11. Multi Parameter Tablet Machine

Testing of five essential physical parameters: hardness, thickness, diameter/length, width and weight of a tablet Equipped with the Auto Alignment device, even the most difficult of tablet shapes are automatically aligned properly to prevent erroneous hardness measurements.

Two hardness measurement principles “Constant Speed” or “Constant Force” with a measuring range of up to 500 or 1000 N can be applied to match existing testers on the market.

Meets the USP chapter 1217 requirements	:	three points calibration and tablet orientation.
---	---	--

Optional Accessories

Balance Mettler, Sartorius or Precisa
Digital caliper

The Q-Doc Quality Documentation Software with a PC. Q-Doc permits storage of a large number of methods as well as test reports. Customized reports are also possible. Q-Doc is a validated Windows 2000/XP/Vista software which fully complies with 21 CFR Part 11. Q-Doc is also available with LIMS capabilities

Specifications

Measurement	:	Constant Speed,
Principles	:	Constant Force
Range	:	0.05-2.5 mm/min 10-99N/sec

Measurement Range	:	
Hardness	:	5 - 500 (+/-1 N)
Accuracy	:	10 -1000 (+/-2 N)
Diameter/Length	:	0.00 - 30.00 mm
Accuracy	:	+/-0.05 mm
Thickness	:	0.00 - 18.00 mm
Accuracy	:	+/-0.05 mm
Weight	:	50 mg - 5 g
Accuracy	:	+/-0.2 mg 5 g - 30 g +/-1 mg
Power	:	230 V/50 Hz

12. Laboratory Filtration Kit

Consisting of:

- (1) All Glass Filter Holder : 1 No.
47mm, made of
Borosilicate glass
- (2) Vacuum/Pressure Pump, : 1 No.
Oil free portable pump,
pressure up to 4 bar :
- (3) Filter Membrane, White: 1 Pkt. each
plain membrane, :
suitable for filtering
organic Solvents, :
diameter-47 mm and
porosity- 0.2µm and
0.45µm :
- (4) Funnel Cover, Silicone : 1 No.
- (5) 33mm PVDF syringe
Filter Unit : 1 No.
- (6) 0.2 mm Vacuum Trap : 1 No.
- (7) Filter Forceps, SS 316
forceps with beveled
and unserrated tips : 1 No.
- (8) Silicone Hose, 1.4 M. : 1 No.
- (9) Filter Membrane, 33 mm,: 1 Pkt.
0.45mm, PVDF,250/Pkt.
- (10) Filter Membrane, 33 mm,: 1 Pkt.
0.22mm, PVDF, 250/Pkt. :

Desk Top Computer

Processor	:	Intel Core i3-3240 or better, (3M Cache, 3,4 GHz.
OS	:	Windows 8, Single Language English
Memory	:	4 GB DDR3

Hard Drive	: 3.5", 1TB, 7200 RPM, SATA
Video Card	: Intel Integrated Graphics
Optical Drive	: HH SATA tray load DVD RW
Monitor	: 18.5" LED Monitor
Warranty	: 3 Year
UPS	: 800 VA

With Multimedia Kit, Card Reader, Keyboard and Mouse, HDMI port, 4-6 USB ports, Wifi Ready

Refrigerator

Capacity	: 310 Litres
Rating	: 4 Star
Warranty of Compressor:	5 Years

Frost free, Double Doors, CFC Free, with Stand and Stabilizer (1KVA)

TENDER SCHEDULE

1. The period for which the rates to be in firm—One year.
2. The rates should be inclusive of sales tax and for free delivery at the department concerned, otherwise tender will be rejected.
3. Advance payments conditions will not be considered.
4. Tender forms are neither transferable nor the cost of it will be refundable under any circumstances.
5. Tenders are to be sent with details and illustrated pamphlet (wherever necessary) neatly typed. Manuscripts should be avoided as far as possible.
6. Tenders should be accompanied with a formal agreement executed in ₹ 100 Kerala non-judicial stamp paper together with required sum of Earnest Money Deposits.
7. Earnest Money Deposits (One percentage) of the total cost of the articles quoted for, can be remitted by way of crossed Demand Drafts drawn in favour of the Principal, Government Medical College, Thiruvananthapuram.
8. Firms will be exempted from furnishing Earnest Money Deposit if they had produced the duly attested copies of relevant certificate along with their tenders.
9. Tenders unsealed, unsigned or submitted without superscription as specified are liable to be rejected.

10. Tenders submitted with insufficient Earnest Money Deposit will not be entertained.
11. Belated tenders will not be considered under any circumstances.
12. For Medicines, Surgicals and Materials, Samples should be submitted along with the tender. Tenders without samples will not be considered for selection.
13. Tenders should bear the Signature and full Postal address of the tenderers wherever necessary.
14. The undersigned reserves the right to reject any tender wholly or partially without any prior information to the tenders.
15. Specifications of the items are given in the list.
16. Negotiation will be made only with the firm offering lowest price whose specifications confirms to the requirements in the tender notice.
17. Tender will be opened in the presence of the such of the tenderers or his authorized representatives who may present at the time. They should have valid authorization letters.
18. More details can be had from this office on working hours.
19. Amount quoted should be in Indian Rupees.

(2)

No. C1/22735(12)/2013/MCT. 6th December 2013.

Sealed competitive tenders in the form prescribed are invited for the supply of the following items at three Pharmacy Colleges attached to Government Medical Colleges at Kottayam, Alappuzha & Kozhikode. Tenders should be superscribed as tender numbered and addressed to the Principal, Medical College, Thiruvananthapuram, Pin-695 011. Details can be had from this office on working hours till 3.30 p. m.

Sl No.	Tender No. and Details of Items	Quantity (Nos.)
1	239/2013—Spectrofluoro Meter	3
2	240/2013—Moisture Balance	3

Cost of tender form (Inclusive of VAT) :

Original— ₹ 3,780 for Sl. No. 1 and ₹ 756 for Sl. No. 2.

Duplicate—₹ 1,890 for Sl. No. 1 and ₹ 378 for Sl. No. 2.

Last date and time for sale of tender form—7-1-2014, 2.30 p. m.

Last date and time of receipt of tender form—8-1-2014, 2.30 p. m.

Date and time for opening of tender—8-1-2014,
3 p. m.

Tender forms can be had from the Office of the Principal, Medical College, Thiruvananthapuram-11 upto 3 p. m. on working days. Tender forms will be sent through post if required provided due amount with sufficient postage charge (₹ 50 each) is sent by way of Money Order in favour of the Principal, Medical College, Thiruvananthapuram, Kerala State, Pin-695 011. Money Order should reach this office on or before 21-12-2013. No other forms of remittance will be accepted.

കോട്ടയം, കോഴിക്കോട്, ആലപ്പുഴ എന്നീ മെഡിക്കൽ കോളേജുകളിലെ ഫാർമസി കോളേജുകളിലേക്ക് ആവശ്യമായ താഴെപ്പറയുന്ന സാധനങ്ങൾ വിതരണം ചെയ്യുന്നതിന് നിശ്ചിത ഫോറത്തിൽ മുദ്രവച്ച ദർഘാസുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു. ദർഘാസുകൾ ദർഘാസ് നമ്പരോടുകൂടി നിശ്ചിത തീയതിക്കുമുമ്പ് പ്രിൻസിപ്പാൾ, മെഡിക്കൽ കോളേജ്, തിരുവനന്തപുരം-695 011 എന്ന വിലാസത്തിൽ എത്തിച്ചിരിക്കേണ്ടതാണ്.

ക്രമ നമ്പർ	ദർഘാസ് നമ്പരും സാധനങ്ങളുടെ പേരും	എണ്ണം
1	239/13—സ്പെട്രോഫ്ളൂറോ മീറ്റർ	3
2	240/13—മോയിസ്ചർ ബാലൻസ്	3

ദർഘാസ് ഫോറത്തിന്റെ വില :

അസ്സൽ—ക്രമ നമ്പർ 1-ന് ₹ 3,780, ക്രമ നമ്പർ 2-ന് ₹ 756.

ശരിപ്പകർപ്പ്—ക്രമ നമ്പർ 1-ന് ₹ 1,890, ക്രമ നമ്പർ 2-ന് ₹ 378.

ദർഘാസ് ഫോറം വിൽക്കുന്ന അവസാന തീയതിയും സമയവും—7-1-2014, 2.30 മണി.

ദർഘാസ് ഫോറം സ്വീകരിക്കുന്ന അവസാന തീയതിയും സമയവും—8-1-2014, 2.30 മണി.

ദർഘാസ് തുറക്കുന്ന തീയതിയും സമയവും—8-1-2014, 3 മണി.

ദർഘാസ് ഫോറങ്ങൾ തപാലിൽ വേണ്ടവർ അയയ്ക്കുന്നതിനുവേണ്ട തുക (₹ 50) അധികമായി അയയ്ക്കേണ്ടതാണ്. ദർഘാസ് ഫോറത്തിന്റെ വില ഓഫീസിൽ നേരിട്ടോ മണിയോർഡറായോ സ്വീകരിക്കുന്നതാണ്. അല്ലാത്തതുള്ള എല്ലാ അപേക്ഷകളും നിരസിക്കുന്നതാണ്. ഫോറത്തോടൊപ്പം മുദ്രപത്രങ്ങളോ മറ്റേതെങ്കിലും ഡോക്യുമെന്റുകളോ അയയ്ക്കാൻ ആവശ്യപ്പെട്ടുകൊണ്ടുള്ളതും മേൽവിലാസം വ്യക്തമല്ലാത്തതുമായ മണിയോർഡറുകൾ നിരസിക്കുന്നതാണ്. മണിയോർഡറുകൾവഴി ഫോറങ്ങൾ ആവശ്യപ്പെട്ടുകൊണ്ടുള്ള അപേക്ഷകൾ ഈ ഓഫീസിൽ 21-12-2013-ന് മുമ്പ് കിട്ടിയിരിക്കണം. താമസിച്ച് കിട്ടുന്ന മണിയോർഡറുകൾ നിരസിക്കുന്നതാണ്. ദർഘാസ് ഫോറം കൈമാറാൻ പാടില്ലാത്തതും ഒരിക്കൽ അടച്ച തുക തിരിച്ചു കിട്ടുന്നതുമല്ല. താമസിച്ച് കിട്ടുന്ന ദർഘാസുകൾ യാതൊരു കാരണവശാലും സ്വീകരിക്കുന്നതല്ല. കൂടുതൽ വിവരങ്ങൾ പ്രവൃത്തി ദിവസങ്ങളിൽ ഈ ഓഫീസിൽനിന്നും ലഭിക്കുന്നതാണ്.

Spectrofluorometer

Light Source	: Xe lamp with shielded lamp house, 150W.
Wavelength range	: Zero order, 200-750 nm.
Wavelength range accuracy	: Should be at least ± 2 nm.
Spectral bandwidth	: Variable and minimum should be 2.5, 5, 10, 20 nm on both excitation and emission monochromators.
W.L. scan speed	: Variable in the range of 20-20,000nm/min for both excitation and emission monochromators.
Sensitivity	: S/N (Signal to Noise) Ratio RMS 1600:1 or better, Raman band of water Peak to Peak using SBW 5 nm on both EX. and EM. Monochromators, 2 sec. response time at 350nm Excitation wavelength.

Resolution : 2.5nm (at 546.1nm)

Automatic Higher order diffraction cut filters to be provided for eliminating higher order peaks.

Automatic sensitivity selection system to be provided as standard.

Automatic Gain control should be provided as standard.

Should have start button for starting measurement immediately after placing the sample for kinetic measurements.

Should have auto accessory recognition software for identifying the accessories and to select necessary parameters.

Should be provided with Mercury Lamp for validation of the instrument.

Should be on field upgradable to accessories like Powder, film holders, Micro volume sample measurement as low as a single drop, fluorescence Anisotropy, etc.,

Should Provide Windows Based Software For Quantitative Measurement, Spectrum Measurement, Time Course Measurement, Fixed Wavelength Measurement, 3d Fluorescence Measurement, Abs Measurement and Spectral Correction, Validation, Spectral Analysis, File Viewer and Facility for Documentation & Presentation.

Should provide 1No. spare Fluorescence cell, PC, Printer, UPS.

Moisture Analyzer

Specifications

Capacity	: 124g.
Readability	: 1mg/0.01%

Repeatability : 1g/0.2% 10g/0.02%
 Linearity : 1.5mg,
 Pan Size : Ø 100/3.9
 Temperature range : 30-230°C

Features

- Choice of heat sources: Halogen, Infrared or Dark Radiator
 - Temperature range 30-230°C/1°C
 - Integrated balance (124g/0.1mg., 124g/1mg.)
 - Interface RS232, PC and Printer.
 - VFD Display and 10 key Keyboard (XM 60)
 - Anti-theft code.
 - Memory Storage of upto 20 methods.
 - Supplied with 30 aluminum dishes.
 - * Variety of start-up and end-point control selections for optimum control.
 - * Language independent prompts operated by symbol keys.
 - * Easy access sample holder.
- Quality Assurance
- * Results printer in GLP guideline format.
 - * Numerical and statistical analysis available.
 - Real-time results shown in display.

TENDER SCHEDULE

1. The period for which the rates to be in firm—One year.
2. The rates should be inclusive of sales taxes, otherwise tender will be rejected without any notice and for free delivery at the department concerned, otherwise the tender will be rejected.
3. Advance payments conditions will not be considered.
4. Tender forms are neither transferable nor the cost of it will be refundable under any circumstances.
5. Tenders are to be sent with details and illustrated pamphlet (wherever necessary) neatly typed. Manuscripts should be avoided as far as possible.
6. Tenders should be accompanied with a formal agreement executed in ₹ 100 Kerala non-judicial Stamp Paper together with required sum of Earnest Money Deposits.
7. Earnest Money Deposit (One percentage) of the total cost of the articles quoted for, can be remitted by way of crossed Demand Drafts drawn in favour of the Principal, Government Medical College, Thiruvananthapuram.
8. Firms will be exempted from furnishing Earnest Money Deposit if they had produced the duly

attested copies of relevant certificate along with their tenders.

9. Tenders unsealed, unsigned or submitted without superscription as specified are liable to be rejected.
10. Tenders submitted with insufficient Earnest Money Deposit will not be entertained.
11. Belated tenders will not be considered under any circumstances.
12. For Medicines, Surgicals and Materials, Samples should be submitted along with the tender. Tenders without samples will not be considered for selection.
13. Tenders should bear the Signature and full Postal address of the tenderers wherever necessary.
14. The undersigned reserves the right to reject any tender wholly or partially without any prior information to the tenderers.
15. Specifications of the items are given in the list.
16. Negotiation will be made only with the firm offering lowest price whose specification confirms to the requirements in the tender notice.
17. Tender will be opened in the presence of such of the tenderers or his authorized representatives who may present at the time. They should have valid authorization letters.
18. More details can be had from this office on working hours.

(3)

No. C1/22735 (14)/2013/MCT. 6th December 2013.

Sealed competitive tenders in the form prescribed are invited for the supply of the following items at Four Pharmacy Colleges attached to Government Medical Colleges at Thiruvananthapuram, Kottayam, Alappuzha & Kozhikode. Tenders should be superscribed as tender numbered and addressed to the Principal, Medical College, Thiruvananthapuram Pin-695 011. Details can be had from this office on working hours till 3.30 p. m.

Sl No.	Tender No. and Details of Items	Quantity Nos.
1	241/2013—Micropipette-0.5 to 10µl.	8
2	242/2013—Micropipette-10 to 100µl.	8
3	243/2013—Micropipette-100 to 1000µl.	8

Cost of Tender form (Inclusive of VAT) :

Original— ₹ 420.

Duplicate— ₹ 210.

Last date and time for sale of tender form—
7-1-2014, 2.30 p. m.

Last date and time of receipt of tender—
8-1-2014, 2.30 p. m.

Date and time for opening of tender—8-1-2014,
3 p. m.

Tender forms can be had from the office the Principal, Medical College, Thiruvananthapuram-11 upto 3 p. m. on working days. Tender forms will be sent through post if required provided due amount with sufficient postage charge (₹ 50 each) is sent by way of Money Order in favour of the Principal, Medical College, Thiruvananthapuram, Kerala State, Pin-695 011. Money Order should reach this office on or before 21-12-2013. No other forms of remittance will be accepted.

തിരുവനന്തപുരം, കോട്ടയം, കോഴിക്കോട്, ആലപ്പുഴ എന്നീ മെഡിക്കൽ കോളേജുകളിലെ 4 ഫാർമസി കോളേജുകളിലേക്ക് ആവശ്യമായ താഴെപ്പറയുന്ന സാധനങ്ങൾ വിതരണം ചെയ്യുന്നതിന് നിശ്ചിത ഫോറത്തിൽ മുദ്രവച്ച ദർഘാസുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു. ദർഘാസുകൾ ദർഘാസ് നമ്പരോടുകൂടി

നിശ്ചിത തീയതിക്കുമുമ്പ് പ്രിൻസിപ്പാൾ, മെഡിക്കൽ കോളേജ്, തിരുവനന്തപുരം-695 011 എന്ന വിലാസത്തിൽ എത്തിച്ചിരിക്കേണ്ടതാണ്.

ക്രമ നമ്പർ	ദർഘാസ് നമ്പരും സാധനങ്ങളുടെ പേരും	എണ്ണം
1	241/13—മൈക്രോപിപ്പറ്റ്-0.5-10µl.	8
2	242/13—മൈക്രോപിപ്പറ്റ്-10-100µl.	8
3	243/13—മൈക്രോപിപ്പറ്റ്-100-1000µl.	8

ദർഘാസ് ഫോറത്തിന്റെ വില (വാറ്റ് ഉൾപ്പെടെ) :

അസ്സൽ—₹ 420.

ശരിപ്പകർപ്പ്—₹ 210.

ദർഘാസ് ഫോറം വിൽക്കുന്ന അവസാന തീയതിയും സമയവും—7-1-2014, 2.30 മണി.

ദർഘാസ് ഫോറം സ്വീകരിക്കുന്ന അവസാന തീയതിയും സമയവും—8-1-2014, 2.30 മണി.

ദർഘാസ് തുറക്കുന്ന തീയതിയും സമയവും—8-1-2014, 3 മണി.

ദർഘാസ് ഫോറങ്ങൾ തപാലിൽ വേണ്ട വർ അയയ്ക്കുന്നതിനുവേണ്ട തുക (₹ 50) അധികമായി അയയ്ക്കേണ്ടതാണ്. ദർഘാസ് ഫോറത്തിന്റെ വില ഓഫീസിൽ നേരിട്ടോ മണിയോർഡറായോ സ്വീകരിക്കുന്നതാണ്. അല്ലാത്തതുള്ള എല്ലാ അപേക്ഷകളും നിരസിക്കുന്നതാണ്. ഫോറത്തോടൊപ്പം മുദ്രപത്രങ്ങളോ മറ്റേതെങ്കിലും ഡോക്യുമെന്റുകളോ അയയ്ക്കാൻ ആവശ്യപ്പെട്ടുകൊണ്ടുള്ളതും മേൽവിലാസം വ്യക്തമല്ലാത്തതുമായ മണിയോർഡറുകൾ നിരസിക്കുന്നതാണ്. മണിയോർഡറുകൾവഴി ഫോറങ്ങൾ ആവശ്യപ്പെട്ടുകൊണ്ടുള്ള അപേക്ഷകൾ ഈ ഓഫീസിൽ 21-12-2013-ന് മുമ്പ് കിട്ടിയിരിക്കണം. താമസിച്ച് കിട്ടുന്ന മണിയോർഡറുകൾ നിരസിക്കുന്നതാണ്. ദർഘാസ് ഫോറം കൈമാറാൻ പാടില്ലാത്തതും ഒരിക്കൽ അടച്ച തുക തിരിച്ചു കിട്ടുന്നതുമല്ല. താമസിച്ച് കിട്ടുന്ന ദർഘാസുകൾ യാതൊരു കാരണവശാലും സ്വീകരിക്കുന്നതല്ല. കൂടുതൽ വിവരങ്ങൾ പ്രവൃത്തി ദിവസങ്ങളിൽ ഈ ഓഫീസിൽനിന്നും ലഭിക്കുന്നതാണ്.

17. Micropipette

- Adjustable volume single channel air displacement pipette with soft touch tip ejector, extremely smooth piston movement, rotating plunger sets volume within range.
- Volume range-(a) 0.5-10µl with 0.1µl increment.
(b) 10-100µl with 1µl increment.
(c) 100-1000µl with 1µl increment.

Volume display : 1n 4 digits

Warranty : 3 years

NABL calibration facility either in Chennai or in Kerala.

Fully Autoclavable type.

TENDER SCHEDULE

- The period for which the rates to be in firm—One year.
- The rates should be inclusive of sales tax and for free delivery at the department concerned, otherwise the tender will be rejected.
- Advance payments conditions will not be considered.
- Tender forms are neither transferable nor the cost of it will be refundable under any circumstances.
- Tenders are to be sent with details and illustrated pamphlet (wherever necessary) neatly typed. Manuscripts should be avoided as far as possible.
- Tenders should be accompanied with a formal agreement executed in ₹ 100 Kerala non-judicial Stamp Paper together with required sum of Earnest Money Deposits.
- Earnest Money Deposit (One percentage) of the total cost of the articles quoted for, can be remitted by way of crossed Demand Drafts drawn in favour of the Principal, Government Medical College, Thiruvananthapuram.
- Firms will be exempted from furnishing Earnest Money Deposit if they had produced the duly attested copies of relevant certificate along with their tenders.
- Tenders unsealed, unsigned or submitted without superscription as specified are liable to be rejected.
- Tenders submitted with insufficient Earnest Money Deposit will not be entertained.
- Belated tenders will not be considered under any circumstances.
- For Medicines, Surgicals and Materials, Samples should be submitted along with the tender. Tenders without samples will not be considered for selection.
- Tenders should bear the Signature and full Postal address of the tenderers wherever necessary.

14. The undersigned reserves the right to reject any tender wholly or partially without any prior information to the tenderers.
15. Specifications of the items are given in the list.
16. Negotiation will be made only with the firm offering lowest price whose specification confirm to the requirements in the tender notice.
17. Tender will be opened in the presence of such of the tenderers or his authorized representatives who may present at the time. They should have valid authorization letters.
18. More details can be had from this office on working hours.

Government Medical College,
Thiruvananthapuram.

(Sd.)
Principal.

ദർഘാസ് പരസ്യം

(1)

നമ്പർ എ6-5899/2013(I)

2013 ഡിസംബർ 5.

തിരുവനന്തപുരം തെക്കാട് സ്ത്രീകളുടെയും കുട്ടികളുടെയും ആശുപത്രിയിലെ വന്ധ്യതാനിവാരണ ക്ലിനിക്കിലേക്ക് ₹ 30,47,000-ന് മരുന്നുകളും ഉപകരണങ്ങളും വിതരണം ചെയ്യുന്നതിന് മുന്നോടിയായി ദർഘാസ് ക്ഷണിച്ചുകൊള്ളുന്നു. (1) സൈമൺ വാഷിംഗ് മീഡിയം (2) സിംഗിൾ ഗ്രേഡിന്റ് (60%) ഹൈ ഡെൻസിറ്റി സൊല്യൂഷൻ ഫോർ സൈമൺ വാഷിംഗ് (3) ഡോണർ സ്ക്രീം (4) ഐ. യു. ഐ. ക്യാനൂല (5) ട്രാൻസ്ഫർ പിപ്പറ്റ് (6) സെൻട്രിഫ്യൂജ് ട്യൂബ് (7) സൈമൺ കളക്ഷൻ ബോട്ടിൽ (8) ക്രയോകാൻ ഫോർ സ്ക്രീം ബാങ്ക് (9) സ്ക്രീം കൗണ്ടിംഗ് ചേംബർ ഫോർ സൈമൺ അനാലിസിസ് ഓരോ ഇനത്തിനുമുള്ള സാങ്കേതിക വിവരണങ്ങൾ എല്ലാ പ്രവൃത്തി ദിവസങ്ങളിൽ ആഫീസിൽ നിന്നും ലഭ്യമാണ്. കൂടാതെ www.dhs.kerala.gov.in എന്ന വെബ്സൈറ്റിലും ലഭ്യമാണ്. അപ്രകാരം നിർദ്ദേശിച്ചിട്ടുള്ള തരത്തിലും, നിലവാരത്തിലുമുള്ള ഇനങ്ങളുടെ വില നിരക്ക് രേഖപ്പെടുത്തിയ ദർഘാസുകൾ മാത്രമേ അംഗീകാരത്തിന് പരിഗണിക്കുകയുള്ളൂ. ദർഘാസ് ഫോറം പ്രവൃത്തി സമയങ്ങളിൽ ആഫീസിൽ നിന്നും 3-1-2014-ന് ഉച്ചയ്ക്ക് 12 മണി വരെ ലഭിക്കുന്നതാണ്. അസ്സൽ ഫോറത്തിനും ഡ്യൂപ്ലിക്കേറ്റിനും ഗവൺമെന്റ് നിശ്ചയിച്ചിട്ടുള്ള തുക+വാറ്റ് ആഫീസിൽ അടച്ച് കൈപ്പറ്റാവുന്നതാണ്. ദർഘാസ് ഉള്ളടക്കം ചെയ്ത കവറിന് പുറത്ത് ദർഘാസ് നമ്പർ A6-5899/2013-14(I) സൂപ്രണ്ട്, സ്ത്രീകളുടെയും കുട്ടികളുടെയും ആശുപത്രി, തെക്കാട്, തിരുവനന്തപുരം എന്ന് രേഖപ്പെടുത്തിയിരിക്കണം.

ദർഘാസുകൾ സമർപ്പിക്കേണ്ട അവസാന തീയതി 3-1-2014 ഉച്ചയ്ക്ക് 1 മണി വരെയാണ്. ദർഘാസുകൾ അന്നേ ദിവസം 2 മണിക്ക് സമർപ്പിച്ചിട്ടുള്ള വ്യക്തികളുടെയോ അവർ നിർദ്ദേശിക്കുന്ന വ്യക്തികളുടെയോ സാന്നിധ്യത്തിൽ തുറക്കുന്നതാണ്. ഓരോ ഇനത്തിന്റെയും ഗുണമേന്മ ക്ലിനിക്കൽ കമ്മിറ്റി മുമ്പാകെ ബോധ്യപ്പെട്ട ശേഷമേ ദർഘാസ് സ്വീകരിക്കുകയുള്ളൂ. അപ്രകാരം ഗുണനിലവാരം

പരിശോധിച്ച് നിർദ്ദേശിക്കപ്പെടുന്ന ഏറ്റവും കുറഞ്ഞ നിരക്കിലുള്ള (ഓരോ ഇനത്തിനും പ്രത്യേകം പ്രത്യേകമായി) ദർഘാസുകൾ മാത്രമേ അംഗീകാരം നൽകി സ്വീകരിക്കുകയുള്ളൂ. ദർഘാസുകൾ സ്വീകരിക്കുന്നതിനും നിരസിക്കുന്നതിനും മുൻപ് പൂർണ്ണാധികാരം സൂപ്രണ്ടിൽ നിക്ഷിപ്തമായിരിക്കും. സമർപ്പിക്കുന്ന ദർഘാസ് ഫോറങ്ങൾ ഈ ആശുപത്രിയിൽ നിന്നും വാങ്ങിയിട്ടുള്ളതും പ്രസ്തുത സ്ഥാപനത്തിലെ മേലധികാരിയുടെ കൈയൊപ്പും ആഫീസ് മുദ്രയും ഉള്ളതായിരിക്കണം.

ദർഘാസുകൾ സമർപ്പിക്കുമ്പോൾ ₹ 100 വിലയുള്ള മുദ്രപത്രത്തിലുള്ള ഉടമ്പടിയും ₹ 5,000 (അയ്യായിരം) നിരതദ്രവ്യമായി സൂപ്രണ്ടിന്റെ പേരിൽ മറാവുന്ന ഡിമാന്റ് ഡ്രാഫ്റ്റും ഉള്ളടക്കം ചെയ്തിരിക്കണം. ദർഘാസുകൾ അംഗീകരിച്ച് ലഭിക്കുന്ന കരാറുകാരൻ നിർദ്ദിഷ്ട സമയപരിധി കുള്ളിൽതന്നെ സാധന സാമഗ്രികൾ തൃപ്തികരമാവിധം വിതരണം ചെയ്യേണ്ടതാണ്.

കരാറുകാരൻ നിബന്ധനകൾക്ക് വിപരീതമായി പ്രവർത്തിച്ചാൽ അപ്രകാരം സംഭവിക്കുന്നതുമൂലമുള്ള നഷ്ടപരിഹാരം കരാറുകാരിൽ നിന്നും ഈടാക്കുന്നതിന് സൂപ്രണ്ടിന് അധികാരം ഉണ്ടായിരിക്കുന്നതാണ്. കൂടുതൽ വിവരങ്ങൾ പ്രവൃത്തി ദിവസങ്ങളിൽ ആഫീസിൽ നിന്നും അറിയാവുന്നതാണ്.

(2)

നമ്പർ എ6-5899/2013(II)

2013 ഡിസംബർ 5.

തിരുവനന്തപുരം തെക്കാട് സ്ത്രീകളുടെയും കുട്ടികളുടെയും ആശുപത്രിയിലെ വന്ധ്യതാനിവാരണ ക്ലിനിക്കിലേക്ക് ₹ 13,35,000-ന് മരുന്നുകളും ഉപകരണങ്ങളും വിതരണം ചെയ്യുന്നതിന് മുന്നോടിയായി ദർഘാസുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു. (1) ലിക്വിഡ് നൈഡ്രജൻ (ലിറ്റർ) (2) കൂക്ക് ഐ. യു. ഐ. കത്തീറ്റർ കൂക്ക് ഷാപ്പാർഡ് ഐ. യു. ഐ. കാനൂല (3) ഫോളീസ് കത്തീറ്റർ (നമ്പർ 8) (4) K JET SIVF 7019 ET കത്തീറ്റർ (5) ഓവും പിക്അപ്പ് നീഡിൽ KOSN 1735 (ഓവും നീഡിൽ) സിങ്കിൾ ലുമെൻ (6) ഓവും പിക്അപ്പ് നീഡിൽ KOSN 1735 (ഓവും നീഡിൽ) ഡബിൾ ലുമെൻ (7) നോവി കോർണുവൽ കാനുലേഷൻ സെറ്റ് (8) സോഫ്റ്റ്വെയർ ഡെവലപ്പ്മെന്റ് ആന്റ് കമ്പ്യൂട്ടറൈസേഷൻ ഓഫ് ഏ വി. എഫ് ക്ലിനിക്.

ഓരോ ഇനത്തിനുമുള്ള സാങ്കേതിക വിവരണങ്ങൾ പ്രവൃത്തി ദിവസങ്ങളിൽ ആഫീസിൽ നിന്നും ലഭ്യമാണ്. കൂടാതെ www.dhs.kerala.gov.in എന്ന വെബ്സൈറ്റിലും ലഭ്യമാണ്. അപ്രകാരം നിർദ്ദേശിച്ചിട്ടുള്ള തരത്തിലും, നിലവാരത്തിലുമുള്ള ഇനങ്ങളുടെ വില നിരക്ക് രേഖപ്പെടുത്തിയ ദർഘാസുകൾ മാത്രമേ അംഗീകാരത്തിന് പരിഗണിക്കുകയുള്ളൂ. ദർഘാസ് ഫോറം പ്രവൃത്തി സമയങ്ങളിൽ ആഫീസിൽ നിന്നും 3-1-2014-ന് ഉച്ചയ്ക്ക് 12 മണി വരെ ലഭിക്കുന്നതാണ്. അസ്സൽ ഫോറത്തിനും ഡ്യൂപ്ലിക്കേറ്റിനും ഗവൺമെന്റ് നിശ്ചയിച്ചിട്ടുള്ള തുക+വാറ്റ് ആഫീസിൽ അടച്ച് കൈപ്പറ്റാവുന്നതാണ്. ദർഘാസ് ഉള്ളടക്കം ചെയ്ത കവറിന് പുറത്ത് ദർഘാസ് നമ്പർ

A6-5899/2013-14(II) സുപ്രണ്ട്, സ്ത്രീകളുടെയും കുട്ടികളുടെയും ആശുപത്രി, തൈക്കാട്, തിരുവനന്തപുരം എന്നും രേഖപ്പെടുത്തിയിരിക്കണം.

ദർഘാസുകൾ സമർപ്പിക്കേണ്ട അവസാന തീയതി 3-1-2014 ഉച്ചയ്ക്ക് 1 മണി വരെയാണ്. ദർഘാസുകൾ അന്നേ ദിവസം പകൽ 2 മണിക്ക് സമർപ്പിച്ചിട്ടുള്ള വ്യക്തികളുടെയോ അവർ നിർദ്ദേശിക്കുന്ന വ്യക്തികളുടെയോ സാന്നിധ്യത്തിൽ തുറക്കുന്നതാണ്. ഓരോ ഇനത്തിന്റേയും ഗുണമേന്മ ടെക്നിക്കൽ കമ്മിറ്റി മുമ്പാകെ ബോധ്യപ്പെട്ട ശേഷമേ ദർഘാസ് സ്ഥിരപ്പെടുത്തുകയുള്ളൂ. അപ്രകാരം ഗുണനിലവാരം പരിശോധിച്ച് നിർദ്ദേശിക്കപ്പെടുന്ന ഏറ്റവും കുറഞ്ഞ നിരക്കിലുള്ള (ഓരോ ഇനത്തിനും പ്രത്യേകം പ്രത്യേകമായി) ദർഘാസുകൾ മാത്രമേ അംഗീകാരം നൽകി സ്ഥിരപ്പെടുത്തുകയുള്ളൂ. ദർഘാസുകൾ സ്വീകരിക്കുന്നതിനും നിരസിക്കുന്നതിനുമുള്ള പൂർണ്ണാധികാരം സുപ്രണ്ടിൽ നിക്ഷിപ്തമായിരിക്കും.

സമർപ്പിക്കുന്ന ദർഘാസ് ഫോറങ്ങൾ ഈ ആശുപത്രിയിൽ നിന്നും വാങ്ങിയിട്ടുള്ളതും പ്രസ്തുത സ്ഥാപനത്തിലെ മേലധികാരിയുടെ കൈയൊപ്പും ആഫീസ് മുദ്രയും ഉള്ളതായിരിക്കണം.

ദർഘാസുകൾ സമർപ്പിക്കുമ്പോൾ ₹ 100 വിലയുള്ള മുദ്രപത്രത്തിലുള്ള ഉടമ്പടിയും ₹ 5,000 (അയ്യായിരം) നിരതദ്രവ്യമായി സുപ്രണ്ടിന്റെ പേരിൽ മാനാവുന്ന ഡിമാന്റ് ഡ്രാഫ്റ്റും ഉള്ളടക്കം ചെയ്തിരിക്കണം. ദർഘാസുകൾ അംഗീകരിച്ച് ലഭിക്കുന്ന കരാറുകാരൻ നിർദ്ദിഷ്ട സമയപരിധിക്കുള്ളിൽതന്നെ സാധനസാമഗ്രികൾ തൃപ്തികരമാംവിധം വിതരണം ചെയ്യേണ്ടതാണ്.

കരാറുകാരൻ നിബന്ധനകൾക്ക് വിപരീതമായി പ്രവർത്തിച്ചാൽ അപ്രകാരം സംഭവിക്കുന്നതുമൂലമുള്ള നഷ്ടപരിഹാരം കരാറുകാരിൽ നിന്നും ഈടാക്കുന്നതിന് സുപ്രണ്ടിന് അധികാരം ഉണ്ടായിരിക്കുന്നതാണ്. കൂടുതൽ വിവരങ്ങൾ പ്രവൃത്തി ദിവസങ്ങളിൽ ആഫീസിൽ നിന്നും ലഭിക്കുന്നതാണ്.

സ്ത്രീകളുടെയും കുട്ടികളുടെയും
ആശുപത്രി, തൈക്കാട്,
തിരുവനന്തപുരം.

(ഒപ്പ്)
സുപ്രണ്ട്.